Statement of Deficiencies	**(X1) Provider/Supplier/CLIA Identification Number**	**(X3) Date Survey Completed**
	883920	08/15/2022
Name of Provider or Supplier	**Street Address, City, State**	
Pediatric Place Of Union City, Llc	1345 Edwards Street, Union City, TN	

For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.

(X4) ID Prefix Tag	**Summary Statement of Deficiencies** (Each deficiency should be preceded by full regulatory or LSC identifying information)
J0042	PHYSICAL PLANT AND ENVIRONMENT 491.6(b) Maintenance: The clinic . . . has a preventive maintenance program to ensure that: (1) All essential mechanical, electrical and patient-care equipment is maintained in safe operating condition; This STANDARD is not met as evidenced by: Based on interview the facility failed to develop and implement a system for preventative maintenance program to ensure all mechanical and electrical patient equipment were safe and functioning properly. The findings included: During the entrance conference when asked for the preventative maintenance information, the Physican provided the clinic OSHA manual, which did not adddress preventative maintenance. During an interview on 8/15/2022 at 10:10 AM, the Office Manager stated there was no inspection records for the clinic equipment and there was no contracted biomedical company that inspected the equipment annually.